

Certificate

Certificate no.

CU 72408676 0001

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte. Ltd.
#02-13/18 29 Woodlands Industrial Park
E1. North Tech Lobby 3
Singapore 757716
Singapore

Report Number:

US24BV7S 002

Client Reference:

Carol Escano

Certification acc. to:

UL 61010-1:2012 R7.19
UL 61010-2-010:2019
UL 61010-2-081:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-010:19
CSA C22.2 No. 61010-2-081:19

Product Information

Certified Product:

Genetic Sequencer

Model Designation:

MiSeq i100; MiSeq i100 Plus

Technical Data:

Rated Voltage: AC 100-240V, 50/60Hz
Rated Power: 300W
Protection Class: I

Remarks:

Appendix:

1, 1-4

Date of issue:

2024-09-17
(yr/mo/day)

