

Certificate



Certificate no.

CU 72140553 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte Ltd
29 Woodlands Industrial Park E1
Singapore 757716
Singapore

Test report no.: USA-JR 31470626 002

Client Reference: Carol Rogers Escano

Tested to:

UL 61010-1:2004 R10.08
CAN/CSA-C22.2 NO. 61010-1-04+GI1 (R2009)
CAN/CSA-C22.2 No. 61010.2.081-04 (R2009)
CAN/CSA-C22.2 No. 61010-2-101-04 (R2009)
IEC 61010-2-081:2001+A1
IEC 61010-2-101:2002

Certified Product: DNA Analyzer

License Fee - Units

Model Designation: DX-100-1001, SY-101-1001

7

Rated Voltage: AC 100-120/200-240V, 50/60Hz

Rated Power: 360VA

Protection Class: I

7

Appendix: 1, 1-3

Licensed Test mark:

Date of Issue

(day/mo/yr)

07/04/2014

