

Date : 07/26/2022

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Chad Cunningham

Re. : CU US + Canada Certificate

Type of Equipment : DNA Analyzer
Model Designation : See Certificate
Certificate No. : CU 72227227 0001
File No. : 31470626 009
Engineer/Contact : Jorge Rubio
Standard(s) : UL 61010-1:2012 R11.18
UL 61010-2-081:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) +
A1
CSA C22.2 No. 61010-2-081:19

Dear Mr. Cunningham,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72227227 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body

Dipl.-Ing. M. Glagla
QA Certification Officer

TUV Rheinland
of North America, Inc.

295 Foster St.
Suite 100
Littleton, MA 01460

Tel 978-266-9500
Fax 978-266-9992
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72227227 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte. Ltd.
#02-13/18 29 Woodlands Industrial Park E1.
Singapore 757716
Singapore

Test report no.: USA-JR 31470626 009

Client Reference: Chad Cunningham

Tested to:

UL 61010-1:2012 R11.18
UL 61010-2-081:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-081:19

Certified Product: DNA Analyzer

License Fee - Units

Model Designation: iScan

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 300W

Protection Class: I

Special Remarks: Replaces Certificate CU72200175.

Appendix: 1, 1-6

Licensed Test mark:



Date of Issue

(day/mo/yr)

26/07/2022