

Date : 04/21/2022

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Carol Escano

Re. : CU US + Canada Certificate

Type of Equipment : Next Generation Sequencer
Model Designation : See Certificate
Certificate No. : CU 72225983 0001
File No. : 31790423 022
Engineer/Contact : Jessamyn Duterte
Standard(s) : UL 61010-1:2012 R7.19
UL 61010-2-010:2019
UL 61010-2-081:2019
UL 61010-2-101:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) +
A1
CSA C22.2 No. 61010-2-010:19
CSA C22.2 No. 61010-2-081:19
CSA C22.2 No. 61010-2-101:19

Dear Ms. Escano,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72225983 0001. It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved unit.

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

If we can be of any further assistance to you, please do not hesitate to contact us.

Sincerely yours,
Certification Body

Dipl.-Ing. M. Glagla
QA Certification Officer

Enclosure

TUV Rheinland
of North America, Inc.

295 Foster St.
Suite 100
Littleton, MA 01460

Tel 978-266-9500
Fax 978-266-9992
Toll Free TUV-RHEINLAND
Mail info@tuv.com
Web www.tuv.com

Member of
TUV Rheinland Group

Certificate



Certificate no.

CU 72225983 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina, Inc.
25861 Industrial Blvd.
Hayward CA 94545
USA

Test report no.: USA-JD 31790423 022

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012 R7.19
UL 61010-2-010:2019
UL 61010-2-081:2019
UL 61010-2-101:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-010:19
CSA C22.2 No. 61010-2-081:19
CSA C22.2 No. 61010-2-101:19

Certified Product: Next Generation Sequencer

License Fee - Units

Model Designation: NovaSeq 6000Dx

Rated Voltage: AC 200-240V, 50/60Hz

Rated Current/Power: 15A (2500W)

Protection Class: I

Appendix: 1, 1-8

Licensed Test mark:



Date of Issue

(day/mo/yr)

20/04/2022

Certificate



Certificate no.

CU 72225983 02

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Venture Manufacturing Services Inc.
1621 Barber Lane
Milpitas CA 95035
USA

Test report no.: USA-JD 31790423 022

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012 R7.19
UL 61010-2-010:2019
UL 61010-2-081:2019
UL 61010-2-101:2019
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
CSA C22.2 No. 61010-2-010:19
CSA C22.2 No. 61010-2-081:19
CSA C22.2 No. 61010-2-101:19

Certified Product: Next Generation Sequencer

License Fee - Units

Additional Manufacturing Plant: see above (K2395446)

Licensed Test mark:



Date of Issue

(day/mo/yr)
20/04/2022